

SAMPLE REQUIREMENTS:

- **RNA Nano Assay:**
 - **Total RNA Assay: 25-500 ng/ul**
 - **mRNA Assay: 25-250ng/ ul**
- **DNA 1000 Assay: 0.1- 50 ng/ul**
- **Small RNA Assay: 50- 2000pg/ul**
- **RNA Pico Assay: 200-5000 pg/ul**
- **High Sensivity DNA Assay: 5-500 pg/ul**

- **Sample volume: 3µl minimum.**
- **Tube: 1,5 or 0,5 ml.**
- **All the tubes must be legible and perfect labeled**

After QC checking the samples after will be stored in the service for a week. If the user wants to get back their samples, please come to pick up them as soon as possible.